

Contents

Introduction 1-1

Platelet Collection Process – Overview	1-2
Extended Life Platelet (ELP™) Procedures	1-2
LeukoReduction System (LRS) Procedures	1-3
LRS Turbo (LRST) Procedures	1-4

Before You Start 2-1

Configuration Options	2-1
Configuring the Platelet Collection Procedure End Point	2-2
Configuring the AC Infusion Rate	2-2
Configuring the Inlet:AC Ratio	2-3
Configuring Total Plasma Removed	2-5
Enabling the LeukoReduction System	2-6
Enabling the Single-Needle Option	2-7
Configuring High Flow	2-7
Configuring Centrifuge Step Down	2-8
Product Sampling	2-8
Net Additional Saline Returned to Donor	2-9
Platelet Collection Fluid Volumes	2-9
Volumes Processed During Rinseback	2-10
Changed vs. Affected Values	2-11
Ranges Allowed for Changes	2-12

Extended Life Platelet (ELP™) Collection

Procedures 3-1

Equipment and Supplies	3-1
Setting Up the Spectra System	3-3

Installing the Return Flow Controller: Single-Needle Procedure Only	3-3
Installing the Filler	3-3
Loading the Disposable Tubing Set	3-3
Placing Tubing on the Front Panel	3-3
Installing the Single-Needle Bag: Single-Needle Procedure Only	3-4
Installing the Channel in the Filler	3-5
Priming the Disposable Tubing Set	3-7
For Dual-Needle Procedure	3-7
For Single-Needle Procedure	3-9
For Dual- and Single-Needle Procedure	3-11
Entering Donor Data	3-13
Connecting the Donor	3-15
For Dual-Needle Procedure	3-15
For Single-Needle Procedure	3-15
Starting the Run	3-16
Establishing the Interface: Dual-Needle Procedure	3-18
Establishing the Interface: Single-Needle Procedure	3-18
Collecting Concurrent Plasma	3-19
Ending the Run	3-20
Starting Rinseback	3-21
Disconnecting the Donor	3-22
Removing the Disposable Tubing Set	3-23

LeukoReduced Platelet Collection Procedures. 4-1

Equipment and Supplies	4-1
Setting Up the Spectra System	4-3

Installing the Return Flow Controller: Single-Needle Procedure Only	4-3
Installing the Filler	4-3
Loading the Disposable Tubing Set	4-4
Placing Tubing on the Front Panel	4-4
Installing the Single-Needle Bag: Single-Needle Procedure Only	4-5
Installing the Channel in the Filler	4-5
Priming the Disposable Tubing Set	4-8
For Dual-Needle Procedure	4-8
For Single-Needle Procedure	4-11
For Dual- and Single-Needle Procedures	4-14
Entering Donor Data	4-15
Connecting the Donor	4-17
For Dual-Needle Procedure	4-17
For Single-Needle Procedure	4-17
Starting the Run	4-18
Establishing the Interface: Dual-Needle Procedure	4-20
Establishing the Interface: Single-Needle Procedure	4-21
Collecting Concurrent Plasma	4-22
Platelet Inventory Recovery (PIR)	4-23
Starting Rinseback	4-25
Disconnecting the Donor	4-27
Removing the Disposable Tubing Set	4-27

Helpful Hints and Recovery Procedures 5-1

Viewing AC Data	5-1
Returning Plasma or Platelet Product to Donor at the End of the Procedure	5-2
Yield Scaling Factor (YSF)	5-3

Manual Air Block Recovery	5-4
For Dual-Needle Platelet Procedures	5-4
For Single-Needle Platelet Procedures	5-5
Manual Platelet Inventory Recovery (PIR) from the LRS Chamber	5-6
Saline Platelet Inventory Recovery (PIR)	5-7
For Dual-Needle Platelet Procedures	5-7
For Single-Needle Platelet Procedures	5-8
Terminating LRS/LRST Procedures Early	5-9
During the Run	5-9
During PIR	5-9
Expelling Air (or Saline) from Collection Bags Before Starting Donor Procedure	5-10
Preparing a Double-Platelet Product	5-11
Calculating Anticoagulant in Collected Products	5-11
Manual Spillover Recovery	5-12
For Dual-Needle Procedures	5-12
For Single-Needle Procedures	5-14
Measuring Platelet and WBC Content Codes	5-17
Decreasing Platelet Activation or “Clumping”	5-18
Optimizing Platelet Concentration	5-19

Alarms and Troubleshooting 6-1

CCM INDICATES NO PLATELETS COLLECTED	6-2
Excessive collect line contamination!	6-3
Measure platelet and white blood cell content of this product!	6-4
Post-count in donor may be less than 100,000 platelets/ μ L	6-4
RETURN FLOW TOO SLOW!	6-5
RETURN PRESSURE TOO LOW!	6-5

Spillover detected. Red blood cells in collect line? 6-6
Spillover detected. Returning collect 6-6
Total plasma collected (collect and plasma bags) exceeds
specified limit 6-6
Index i-1